



# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## *Certificate of Accreditation*

*Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:*

### ***Accuracy Calibration Services S. de R.L. de C.V.***

*Juan de Grijalva #07, Fracc. Conquistadores  
Ciudad Rio Bravo, Tamaulipas, México. C.P. 88993*

*(Hereinafter called the Organization) and hereby declares that Organization is accredited  
in accordance with the recognized International Standard:*

**ISO/IEC 17025:2017**

This accreditation demonstrates technical competence for a defined scope and the  
operation of a laboratory quality management system  
(as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

***Dimensional Inspection***  
***(As detailed in the supplement)***

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen  
President

*Initial Accreditation Date:*

June 28, 2023

*Issue Date:*

June 28, 2023

*Expiration Date:*

September 30, 2025

*Accreditation No.:*

69876

*Certificate No.:*

L23-503

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
755 W. Big Beaver, Suite 1325  
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based on a  
continuous accreditation cycle. The validity of this certificate should be  
confirmed through the PJLA website: [www.pjllabs.com](http://www.pjllabs.com)*



# Certificate of Accreditation: Supplement

## Accuracy Calibration Services S. de R.L. de C.V.

Juan de Grijalva # 7, Frac. Conquistadores

Rio Bravo, Tamaulipas, México. C.P. 88993

Contact Name: Hector Millan Espinoza Phone: 899-933-8666

*Accreditation is granted to the facility to perform the following testing:*

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Dimensional Inspection <sup>F</sup>	Industrial Plastic Metal Parts	Geometrical and Dimensional Tolerances GD&T	ASME Y14.5 CMM Mitutoyo 196-444	500 mm x 580 mm x 320 mm (Res.= 0.001 mm)
			ASME Y14.5 Vision System Machine Micro-Vu Mod. M436	180 mm x 180 mm (Res.= 0.000 1 mm)
			ASME Y14.5 Micrometer	Up to 25.4 mm (Res = 0.001 mm)
			ASME Y14.5 Caliper	Up to 150 mm (Res = 0.01 mm)

1. The presence of a superscript F means that the laboratory performs testing of the indicated parameter at its fixed location. Example: Outside Micrometer<sup>F</sup> would mean that the laboratory performs this testing at its fixed location.

